



Nitrile examination gloves

Name:	Master Glove Nitrile PF Simple
Catalogue number:	T.RED.NB21.BL.OL/IT_S
Size:	L
Glove type:	diagnostic, protective, non-sterile
Material:	nitrile rubber (NBR)
Powder content:	none
Surface:	textured fingertips
Cuff finish:	rolled
Shape:	universal, fitting both hands
Color:	blue
AQL:	1.5
Personal Protective Equipment:	category III
Conformity Assessment:	CE Declaration of Conformity
Quantity in unit packaging:	100 pcs.



Qualitative parameters:

The Master Glove gloves meet the requirements of: Regulation (EU) 2017/745 (MDR) on medical devices and Regulation (EU) 2016/425 on Personal Protective Equipment, as well as the standards: EN 455-1, EN 455-2, EN 455-3, EN 455-4, EN 1041:2008, EN 15223-1:2016, EN 14971:2019, EN ISO 10993-1:2018, EN ISO 10993-10:2010, EN ISO 10993-11:2017, EN ISO 374-1, EN ISO 374-2, EN ISO 374-4, EN ISO 374-5, EN ISO 16523-1, EN 388, EN ISO 21420, ASTM D6978-05, ASTM F1671-13, ASTM F1670, ASTM D3577, ASTM D5712, ASTM D6319-19, ASTM F739, EN ISO 16604. The gloves are intended for contact with food in accordance with the requirements of Regulation (EU) 10/2011 and standard EN 1186-1.



Manufacturer:
Shandong Intco Medical Products Co., Ltd.
Qiwang Road No. 9888, Naoshan Industrial
Park, Qingzhou, Shandong, China
www.intcomedical.com



Importer:
SORIMEX sp. z o.o. sp. k.
Równinna 25, 87-100 Toruń, Poland
www.sorimex.eu